

Proficiency Testing Program
Organised by
Center for Laboratory Proficiency Testing
Department of Science Service,
Ministry of Science and Technology, Bangkok, Thailand
Total Suspended Solids (TSS) in water

PTEN - W04 - 1801

16 May 2018

SUMMARY SHEET

Laboratory Name : Technical Training and Support Service, Center for Scientific and Technological Equipment, Walailak University

Laboratory Code : W04 - 1801 - 233

Sample	Number of results	Assigned value (x_{pt}) (mg/L)	Standard deviation for proficiency assessment (σ_{pt}) (mg/L)	Lab's result (mg/L)		z-score
				Value	MU	
A	293	85.6	6.8	86	1.49	0.06
B	293	432.5	34.6	434	7.73	0.04

Note

1. Assigned values (x_{pt}) are the robust averages of the participant results in this round, calculated by using Algorithm A (ISO 13528:2015).
2. Standard deviations for proficiency assessment (σ_{pt}) are the target standard deviations from previous rounds and expert judgement.
3. Each z-score marked with "A" is an outlier ($|z| \geq 3.0$). Therefore, laboratory which is marked with "A" shall investigate the result. Laboratory is also encouraged to review the result which has an absolute z-score value between two and three ($2.0 < |z| < 3.0$) marked with "W".
4. This summary sheet shall be read in conjunction with the final report.

Reported by

H. Suprawee

(Miss Suprawee Wongsuchoto)
Program Coordinator

Approved by

[Signature]

(Mrs. Pochaman Tagheen)
Head of Center for Laboratory Proficiency Testing
Department of Science Service